

Thermo Fisher Scientific hereby certifies that the product identified below is produced, inspected and found to be in compliance with product and quality specification requirements as documented in our ISO 13485:2003 Quality Management System (QMI-SAI Global File No. 1616978) in Monterrey, Mexico.



Janet Cosgrove
Raw Materials and
Product Compliance Manager

The following information represents Product Certification for: Item#: **5970-0520**

Certificate issued: **07/22/2016**

Description: **RACK TT UNWIRE RED; 20MM**

Lot#: **1175379**

Manufactured: **04/13/2016**

Part Number	Description	Common Name	DMF#	Cytotoxicity	USP Class VI	FDA Compliance - 21 CFR
1-2503-45P	RACK,TT UNWIRE, RESMER RED; 20 MM.	COMPONENT PART				
8-0025-15P	RESMER, 20%GF, RED	COMPONENT PART				
8-0025-01	RESIN, RESMER, 20%GF, NAT	SYNDIOTACTIC PS, NATURAL	N/A	N/A	N/A	N/A
8-0098-13	COLOR CONCENTRATE,RED,QUESTRA	RED COLOR CONCENTRATE	N/A	N/A	N/A	N/A

If N/A appears in any of the columns above it means the information is not available. Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity, USP Class VI, and FDA Compliance Information columns.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.