

Certificate of Analysis

Product Name: Architect Magnesium
OEM Part Number: 03P680031
OEM Lot Number: 10022UN16
Storage: 2 - 8°C
Qty Manufactured: 398 Kits

Abbott List Number: 03P680031
Abbott Commodity Number: 39-6927
Expiration Date: 2017-11-25
Package Insert No./Rev. 305531 / R02
Test Date 2016/11/28

Description	Lot Number	Expiration Date
Magnesium R1	10022UN16	2017-11-25
Magnesium R2	10022UN16	2017-11-25

Specification	Specification Limit	Result
Barcode Scan	Acceptable verification on c8000	Complies Pass
Bio-Rad L1	1.80 - 1.99 mg/dL	1.85 Pass
Bio-Rad L1	Lot Number	15651
Bio-Rad L2	4.15 - 4.59 mg/dL	4.29 Pass
Bio-Rad L2	Lot Number	15652
Standard High: Difference	-10 - 10 %	3 Pass
Microbial Content: R1 Beg	<30 CFU/mL	0 Pass
Microbial Content: R1 End	<30 CFU/mL	0 Pass
Microbial Content: R2 Beg	<30 CFU/mL	0 Pass
Microbial Content: R2 End	<30 CFU/mL	0 Pass

For in-vitro diagnostic use.

This product contains animal source material. The bovine material used in the manufacture of this product is derived from bovine blood collected from U.S. sourced cattle slaughtered at a USDA licensed establishment.

This lot meets all current release specifications per quality document MS.MASTER342.

The above product has been manufactured in accordance with Quality Systems Regulations, passed all Thermo Fisher final release specifications, and has met all Abbott acceptance requirements.



Quality Representative

Date: 2016-11-29