



# Certificate of Analysis

## Novex® 10% Tris–Glycine Midi Gel

Product No. WT0102BX10  
Lot No. 12120384  
Date of Manufacture 03DEC12  
Expiration Date 25MAR13

---

### Testing Conditions

Gels were run using Novex® Tris–Glycine SDS Running Buffer and unstained protein markers. The gels were electrophoresed at 225V until the dye front reached the bottom of the gel.

### Migration

Reduced Myosin migrated to  $0.235 \pm 0.050$  Rf, reduced BSA to  $0.500 \pm 0.050$  Rf, and reduced Carbonic Anhydrase to  $0.915 \pm 0.050$  Rf.

### Straightness

Across the gel, the migration of a given protein did not vary more than 5% of the length of the gel.

### Curvature

In each of the two outermost lanes, the migration of a given protein did not vary more than 2% of the length of the gel.

### Appearance

Gels run were free of swirls, bubbles, and debris. Bands were sharp and flat.

### Overall Result

PASS

---

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1–800–955–6288 (US and Canada) or 1–760–603–7200, x2 (all other countries).

Life Technologies  
5791 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.lifetechnologies.com](http://www.lifetechnologies.com)  
For inquiries, contact us at [cofarequests@lifetech.com](mailto:cofarequests@lifetech.com)

---

Rick Fletcher  
Director, Quality  
Issued on 04 Dec 2012