

Certificate of Analysis

Product Name: Cortisol Microparticle Reagent
OEM Part Number: 206751
OEM Lot Number: 33233UN17
Storage: 2 - 8°C

Abbott List Number: 206751
Abbott Commodity Number: 8D15G
Expiration Date: 2017-07-16
Package Insert No./Rev. N/A / N/A
Test Date 2017-04-17

Specification	Specification Limit	Result
Mean Cal A RLU	1200000 - 2000000 RLU	1267219 Pass
B/A RLU Ratio	0.55 - 0.75	0.65 Pass
F/A RLU Ratio	<=0.1	0.1 Pass
Low Control	Lower Limit	2.7 µg/dL
Low Control	Upper Limit	4.6 µg/dL
Low Control	Result	3.9 µg/dL Pass
Low Control	Lot Number	40901
Mid Control	Lower Limit	15.6 µg/dL
Mid Control	Upper Limit	21.1 µg/dL
Mid Control	Result	18.9 µg/dL Pass
Mid Control	Lot Number	40902
High Control	Lower Limit	22.9 µg/dL
High Control	Upper Limit	31.0 µg/dL
High Control	Result	26.8 µg/dL Pass
High Control	Lot number	40903
Calibrator	Lot Number	743144-743149
Conjugate	Lot number	33232UN17
Quantity Manufactured	Report	4 L

For in-vitro diagnostic use.

Cortisol Reagents contain animal source material.

This lot meets all current release specifications per quality document MS.MASTER342. The above product has been manufactured in accordance with Fisher Diagnostics' operating procedure.



Quality Representative

Date: 2017-05-12