

Certificate of Analysis

SequalPrep™ Normalization Plate (96) Kit

Product No. A1051001

Lot No. 1829180

Each lot of SequalPrep™ Normalization Plate (96) Kit is functionally qualified and must meet the following specifications.

At least 95% of the SequalPrep™ Normalization Plate must yield greater than or equal to 16ng of DNA.

No well must yield less than or equal to 5 ng of DNA.

The fold difference between the average of the 250ng samples and the 2000ng samples should be less than 4-fold.

Result: Meets specifications.

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1-800-955-6288 (US and Canada) or 1-760-603-7200, x2 (all other countries).

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com



Leon Lambry
Quality Assurance Senior Manager
Issued on 15 May 2017