

CERTIFICATE OF ANALYSIS

26619 **PageRuler Plus Prestained Protein Ladder**

Packaging Lot: 2873341

Expiry Date: 31.03.2026 (DD.MM.YYYY)

Storage: at -20±5°C

Filling lots for components in package:

| Lot | Quantity | Description |
|---------|------------|--|
| 2847000 | 2 x 250 µL | PageRuler Plus Prestained Protein Ladder |

QUALITY CONTROL

| Parameter | Method | Requirement | Result |
|---------------------------|---|---|----------|
| Appearance | Visual inspection. | Dark blue solution | Conforms |
| Band Pattern and mobility | Separation by 4-20% SDS-PAGE and postelectrophoretic in-gel analysis of band migration. | 9 bands visible with migration variation of each bands in the range of 5-10% from those reported in the product specifications. | Conforms |

ISO CERTIFICATION

Manufactured by Thermo Fisher Scientific Baltics UAB, in compliance with ISO 9001 and ISO 13485 certified quality management system.

Quality authorized by QC: **J. Žilinskienė**

